

# AS-A

## PSA Oxygen Generator Parts List

(All Serial Numbers)



AS100-1 (120 V ~) and AS100-2 (220 V ~)			
	Part Number	Description	Quantity
8,000 Hour/1-Year Service Parts	FI019-1	Coalescing Element (MTP-95-548)	1
	FI019-1	Coalescing Element (MTP-95-548)	2
16,000 Hour/2-Year Service Parts	VA088-1	Equalization Valve Rebuild Kit (304817N)	2
	VA089-1	Feed and Waste Rebuild Kit (304817)	4
	VA495-2	1/8" Drain Valve (SC8256A004V)	1
	VA495-3	1/8" Product Solenoid Valve (SC8256B045V)	1
Associated Items	CB144-1	Circuit Board	
	EL001-1	Electrical, EMI Filter	
	FU014-1	1.0 Amp Fuse	
	FU024-1	.5 Amp Fuse	
	GA001-2	Gauge (0 – 100 psig)	
	GA001-3	Gauge (0 – 200 psig)	
	MU002-1	1/4" Muffler	
	PS211-1	MagnaLube	
	SI001-3*	Sieve w/297.62 lbs./barrel	
	SI004-1*	Sieve w/25 lbs./bucket	
	SW003-3	Pressure Switch	
	SW004-1	Rocker Switch (Green)	
	SW005-1	Rocker Switch (Amber)	
	VA010-1	1/4" Solenoid Valve (Feed and Waste)	
	VA010-2	1/4" Solenoid Valve (O2 Cleaned Equalization)	
VA044-1	1/4" Check Valve		
VA110-1	Equalization, Feed, Waste, Valve Coil (218410-001)		
Supplemental Equipment	AN005-1	Oxygen Analyzer (Maxtec Max O <sub>2</sub> Plus)	
	AN021-1	Oxygen Analyzer (Maxtec Handi)	
	AN075-1	Oxygen Analyzer/Sensor (Maxtec Max O <sub>2</sub> Plus)	

\*The AS-A requires approximately 28 pounds of molecular sieve.



© 2017 AirSep Corporation. All Rights Reserved.  
 AirSep reserves the right to discontinue its products, or change the prices, materials, equipment, quality, descriptions, specifications and/or processes to its products at any time without prior notice and with no further obligation or consequence. Any rights expressly stated herein are reserved by us, as applicable.

**Ordering Information:**  
 260 Creekside Drive  
 Buffalo, NY 14228-2075 U.S.A.  
 Tel: (716) 691-0202 ■ Fax: (716) 691-1255  
 www.airsepcpd.com ■ info@airsep.com

AS-SPL-1 A