

# AS-L PSA Oxygen Generator Parts List

(Serial Number 64266 and Higher)



AS107-1 (120 V ~) and AS107-2 (220 V ~)			
	Part Number	Description	Quantity
8,000 Hour/1-Year Service Parts	FI017-1	Coalescing Element (MTP-95-559)	1
	FI030-1	Pre-Filter Element (FRP-95-209)	2
16,000 Hour/2-Year Service Parts	FI017-1	Coalescing Element (MTP-95-559)	2
	FI030-1	Pre-Filter Element (FRP-95-209)	4
	VA104-1	Feed Valve Rebuild Kit (302283)	2
	VA435-1	Equalization Valve Rebuild Kit (C131206)	1
	VA451-1	Waste Valve Rebuild Kit (C131208)	2
	VA495-2	1/8" Drain Valve (SC8256A004V)	1
Associated Items	CM081-2	Omron PLC (Requires SN for Program)	
	FI103-1	Bacteria Filter Element	
	FI127-1	Carbon Filter Element	
	FU004-1	3.0 Amp Fuse	
	FU014-1	1.0 Amp Fuse	
	FU015-1	2.0 Amp Fuse	
	FU024-1	.5 Amp Fuse	
	GA003-1	Gauge (0 – 100 psig)	
	IL057-3	LED Lamp (White)	
	MU030-1	2" Muffler	
	MU052-1	2" Muffler	
	MU057-1	Muffler Element	
	PS211-1	MagnaLube	
	R0878	Wilkerson R31 (Rebuild Kit Seal)	
	R0879	Wilkerson R31 (O-Ring Kit)	
	RG008-1	1" Regulator Wilkerson R31	
	SI001-3*	Sieve w/297.62 lbs./barrel	
	SI004-1*	Sieve w/25 lbs.	
	VA027-3	3/4" Check Valve	
	VA252-2	1 1/2" Solenoid Valve (Feed)	
VA360-1	Feed Valve Coil (272612-032)		
VA429-1	1 1/2" Valve Operator (Waste)		
VA430-1	1/8" Solenoid Valve for Air Operator (Equalization and Waste)		
VA431-1	1 1/4" Valve Operator (Equalization)		
Supplemental Equipment	AN005-1	Oxygen Analyzer (Maxtec Max O <sub>2</sub> Plus)	
	AN021-1	Oxygen Analyzer (Maxtec Handi)	
	AN075-1	Oxygen Analyzer/Sensor (Maxtec Max O <sub>2</sub> Plus)	

\*The AS-L requires approximately 1,800 pounds of molecular sieve.



© 2017 AirSep Corporation. All Rights Reserved.  
AirSep reserves the right to discontinue its products, or change the prices, materials, equipment, quality, descriptions, specifications and/or processes to its products at any time without prior notice and with no further obligation or consequence. Any rights expressly stated herein are reserved by us, as applicable.

**Ordering Information:**  
260 Creekside Drive  
Buffalo, NY 14228-2075 U.S.A.  
Tel: (716) 691-0202 ■ Fax: (716) 691-1255  
www.airsepcpd.com ■ info@airsep.com  
AS-SPL-19 A